



**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Fırızköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE




**TEST REPORT
DENEY RAPORU**

AB-0583-T
20039115
10-20

Customer name: ROSERV PHARMA SRL
Address: 21, Marea Unire Bvd, Building U6, Room 21, Galat., 800329- Galati County, ROMANIA
Buyer name: -
Contact Person: DANA SAVU
Order No: -
Article No: -
Name and identity of test item: Light blue non-woven medical mask.
The date of receipt of test item: 20.10.2020
Re-submitted/re-confirmation date: -
Date of test: 20.10.2020-28.10.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 5

The Turkish Accreditation Agency (TURKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports.

EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory.
The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.



Date
28.10.2020

Customer Representative
Tuğba AKTAŞ

Head of Testing Laboratory
Sevim A. RAZAK
28.10.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Test reports without signature and seal are not valid.

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T

20039115

10-20

REQUIRED TESTS	RESULT	COMMENTS
MICROBIOLOGICAL TESTS		
Bacterial Filtration Efficiency-BFE	P	TYPE IIR
Microbial Cleanliness(Bioburden)	P	
PHYSICAL PROPERTIES		
Breathability(Differential Pressure)	P	
Splash Resistance	P	
P: Pass F: Fail R: Refer to retailer technologist. Tests results were evaluated according to EN 14683:2019+AC :2019 Tablo 1 limit values.		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified.If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95 %. Tests marked (*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

Gen.1136-2/03

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T

20039115

10-20

TEST RESULTS

BACTERIAL FILTRATION EFFICIENCY (BFE)

Test Metod: (Bacterial Filtration Efficiency Testing –BFE /Ref: EN 14683:2019+AC:2019 Medical Face Masks, Requirements and Test Methods

A specimen of the mask material is clamped between an impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Test Flow Rate	28,3 L/min
Test Flow Time	2 minute
Sample Sizes	5 pieces mask
Microorganism	<i>Staphylococcus aureus</i> ATCC 6538
Bacterial concentration (cfu/ ml)	5x10 ⁵ cfu/ ml
incubation conditions	24 hour, 35°C ± 2°C
Positive control sample average of number of Bacteria (C)	2.85x10 ³ cfu/ ml

RESULTS			Requirement
Number of Test Sample	Test Sample (T) Number of Bacteria (cfu/ml)	Bacterial Filtration Efficiency (% B)	BFE (%)
1	54	%98,1	Type I ≥95 Type II ≥98
2	41	%98,6	
3	48	%98,3	
4	43	%98,5	
5	55	%98,1	

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T

20039115

10-20

TEST RESULTS

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: EN ISO 11737-1:2018

The sample is put in extraciton liquid after shaking well, inoculated on the agar.

After incubation at 30 ± 1 ° C for 72 hours, growth microorganisms are counted on the agar.

	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/g)	16 cfu/g	≤ 30 cfu/g Type I and Type II mask

*cfu= Colony forming unit.

SPLASH RESİSTANCE (ONLY FOR TYPE IIR)

Test Metod: EN 14683:2019+AC :2019 (Clause 5.2.4) the resistance of the medical face mask to penetration of splashes of liquid shall conform to the minimum value given for Type IIR in Table 1

ISO 22609 :2004 Clothing for protection against infectious agents — Medical face masks — Test method for resistance against penetration by synthetic blood (fixed volume, horizontally projected)

Test Condition (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs

6 different samples were taken

	<u>SPLASH RESİSTANCE PRESSURE (kPa)</u>	<u>RESULTS</u>	<u>REQUIREMENT</u>
1	>21.3 kPa	PASS	≥ 16 kPa
2	>21.3 kPa	PASS	
3	>21.3 kPa	PASS	
4	>21.3 kPa	PASS	
5	>21.3 kPa	PASS	
6	>21.3 kPa	PASS	
Average Result	>21.3 kPa	PASS	

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T

20039115

10-20

TEST SONUÇLARI

BREATHABILITY (Differential Pressure)

Test Method: EN 14683:2019+AC :2019 (TS EN 14683+AC:2019) Annex-C

Test Condition (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs

Test area is 25 mm in diameter , 5 different sample was taken

Adjusted airflow is 8 l/min. The differential pressure is read directly using a differential pressure manometer .

SAMPLE	DIFFERENTIAL PRESSURE RESULT	REQUIREMENT
1	37,7 Pa/cm ²	< 60 Pa/cm ²
2	45,0 Pa/cm ²	
3	51,2 Pa/cm ²	
4	40,7 Pa/cm ²	
5	41,5 Pa/cm ²	
Average Result	43,2 Pa/cm ²	